

सी इ पी सी आई प्रयोगशाला
एवं अनुसंधान संस्थान

CEPCI LABORATORY &
RESEARCH INSTITUTE



काश्यु भवन्, मुन्टक्कल्, कोल्लम- 691001, केरल, भारत. CIN:U74999KL1955NPL000300
Cashew Bhavan, Mundakkal, Kollam - 691 001, Kerala, India Tel: +91 474 2761003, 2742704
Fax: +91 474 2742704, Email: ceplab@cashewindia.org Web: www.ceplab.org

TEST CERTIFICATE
ORIGINAL

Date: 09 Nov 2022

ULR - TC608822000033994F

No : 4605

CUSTOMER ADDRESS	SAMPLE DETAILS
Name & Address : Indian Public School Kizhavor, Mukhathala P.O., Kollam-691577 Kollam	Reg.No : CEPC/3518/22-23 Customer Reference : Letter Dated 04.11.2022 Sampling details : Direct Identification Code : Type of Sample : Well water Sample Description : Colourless liquid with tiny suspended particles Condition of the sample : Satisfactory Date of Receipt : 04 Nov 2022 Date of Starting : 04 Nov 2022 Date of Completion : 09 Nov 2022 Condition of packing : Unsealed plastic can
Other details if any :	

Table 1. Organoleptic and physical Parameters

Sl.No	Parameters	Results	Test Methods	Acceptable limit:Drinking water specification- IS 10500 : 2012
I	Colour ,Hazan Unit	Less than 1	IS 3025 (part 4):2002	5
II	Odour	Agreeable	IS 3025 (part 5):2002	Agreeable
III	Total dissolved solids ,mg/l	98	IS 3025 (part 16):2002	500

Page 1 of 2

Haseena
Authorized Signatory
(Chemical)
HASEENA. M
Senior Scientist



Mayarani C.B
Authorized Signatory
(Microbiology)
MAYARANI.C.B
Assistant Principal Scientist